

SEAL-TIGHT ADAPTER SELECTION GUIDE

ADEC STYLE



If syringe has a tip insertion fitting like picture #1, you need Seal-Tight Adapter No. 77021



If syringe has a tip insertion fitting like picture #2, you need Seal-Tight Adapter No. 77021

PRESS RING STYLE



If syringe has a tip insertion fitting like picture #3, you need Seal-Tight Adapter No. 77025



If syringe has a tip insertion fitting like picture #4, you need Seal-Tight Adapter No. 77024



Kerr TotalCare
1717 West Collins Avenue
Orange, CA 92867 USA

Phone- 800.841.1428
Fax- 800.638.7390
KerrTotalCare.com

After verifying your brand/style of syringe, please note the quantity you require below.
Please complete all requested information and fax to 800-638-7390.

Quantity - 77021_____ 77025_____ 77024_____

Name/Office _____

Key Contact(s) _____

Street _____

City _____ State _____ Zip _____

Phone (_____) _____ Fax (_____) _____

ADDITIONAL GUIDELINES FOR SEAL-TIGHT ADAPTERS

77021 –“A”– Adec Style: This adapter will fit ALL Adec Style syringes as referenced on the Adapter Selection Guide, regardless of whether they have been previously converted with Sani-Tip or Pro-Tip adapters. The key to recognizing if the syringe is an Adec Style, is after the original fitting or the Sani-Tip or Pro-Tip adapter is removed, the size of the hole in the face of the syringe is the diameter of a pencil or approximately $5/16$ of an inch in diameter.

(Includes: Adec, Ampco, Aseptico, Beaverstate, Belmont, DentalEZ, Dentrex, Empire, Forest, Hampton, Healthco, MDT, Parkell, Robbins, Schein Euro-C, Scope, Unitek, Vacudent, Viking)

77024 – STANDARD “P”– Press Ring Style: This adapter, like the “A”, must be screwed into the syringe. This MUST be used when the syringe has been previously converted to Sani-Tip or Pro-Tip, since the original Press Ring mechanism is usually gone from the office. The key to recognizing if the syringe is a Press Ring Style, is after the fitting is removed, the size of the hole in the face of the syringe is the diameter of a dime or approximately $1\frac{1}{16}$ of an inch in diameter.
(Includes: Belmont, Biotec, DCI, Dentech, Kavo, Marus, Pelton & Crane, Planmeca, Proma)

77025 – QUICK CONNECT “P”– Press Ring Style: This is to be used on any Press Ring Style syringe that has not previously been converted in any way, and the original Press Ring mechanism is intact.

NOTE: There are two other adapters that are rarely used, the “D” (Dental-Ez) and “M” (Engle-Marco). Both of these syringes will be marked with the logo of the company, and if you suspect that these are the syringes that your office has, please call Customer Care at 800-841-1428.